

## ANALYTICAL REPORT

Eurofins Midland  
1211 W. Florida Ave  
Midland, TX 79701  
Tel: (432)704-5440

Laboratory Job ID: 880-17149-1

Client Project/Site: Ector County Utility District

For:

Parkhill Smith & Cooper Inc.  
4222 85th Street  
Lubbock, Texas 79423

Attn: David Clausen



Authorized for release by:  
7/25/2022 5:08:07 PM

John Builes, Project Manager  
(561)558-4549

[John.Builes@et.eurofinsus.com](mailto:John.Builes@et.eurofinsus.com)

### LINKS

Review your project  
results through



Have a Question?



Visit us at:

[www.eurofinsus.com/Env](http://www.eurofinsus.com/Env)

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Client Sample Results . . . . .	5
QC Sample Results . . . . .	7
QC Association Summary . . . . .	8
Lab Chronicle . . . . .	9
Certification Summary . . . . .	11
Method Summary . . . . .	12
Sample Summary . . . . .	13
Chain of Custody . . . . .	14
Receipt Checklists . . . . .	15

# Definitions/Glossary

Client: Parkhill Smith & Cooper Inc.  
Project/Site: Ector County Utility District

Job ID: 880-17149-1

## Qualifiers

### Metals

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
U	Indicates the analyte was analyzed for but not detected.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: Parkhill Smith & Cooper Inc.  
Project/Site: Ector County Utility District

Job ID: 880-17149-1

---

**Job ID: 880-17149-1**

---

**Laboratory: Eurofins Midland**

---

**Narrative**

**Job Narrative**  
**880-17149-1**

**Receipt**

The samples were received on 7/20/2022 1:20 PM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice.

**Metals**

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

# Client Sample Results

Client: Parkhill Smith & Cooper Inc.  
 Project/Site: Ector County Utility District

Job ID: 880-17149-1

**Client Sample ID: 6033 W. Nixon**

**Lab Sample ID: 880-17149-1**

Date Collected: 07/20/22 07:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0337		0.00400	mg/L		07/22/22 11:45	07/23/22 08:44	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 08:44	1

**Client Sample ID: 1735 Jaynita Ave**

**Lab Sample ID: 880-17149-2**

Date Collected: 07/20/22 07:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0326		0.00400	mg/L		07/22/22 11:45	07/23/22 08:46	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 08:46	1

**Client Sample ID: W. Swan Rd**

**Lab Sample ID: 880-17149-3**

Date Collected: 07/19/22 16:30

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0317		0.00400	mg/L		07/22/22 11:45	07/23/22 08:57	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 08:57	1

**Client Sample ID: 5725 W. Turner St.**

**Lab Sample ID: 880-17149-4**

Date Collected: 07/19/22 11:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0798		0.00400	mg/L		07/22/22 11:45	07/23/22 08:59	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 08:59	1

**Client Sample ID: 10464 W. 13th**

**Lab Sample ID: 880-17149-5**

Date Collected: 07/19/22 06:50

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0325		0.00400	mg/L		07/22/22 11:45	07/23/22 09:02	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 09:02	1

**Client Sample ID: 713 S. Warbonnet Ave.**

**Lab Sample ID: 880-17149-6**

Date Collected: 07/19/22 12:30

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0237		0.00400	mg/L		07/22/22 11:45	07/23/22 09:04	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 09:04	1

# Client Sample Results

Client: Parkhill Smith & Cooper Inc.  
 Project/Site: Ector County Utility District

Job ID: 880-17149-1

**Client Sample ID: 12371 W. April St.**

**Lab Sample ID: 880-17149-7**

Date Collected: 07/19/22 09:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0251		0.00400	mg/L		07/22/22 11:45	07/23/22 09:07	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 09:07	1

**Client Sample ID: 7172 W. Verita St.**

**Lab Sample ID: 880-17149-8**

Date Collected: 07/19/22 09:30

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0286		0.00400	mg/L		07/22/22 11:45	07/23/22 09:09	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 09:09	1

**Client Sample ID: 1039 W. Moss Ave**

**Lab Sample ID: 880-17149-9**

Date Collected: 07/20/22 11:35

Matrix: Drinking Water

Date Received: 07/20/22 13:20

**Method: 200.8 - Metals (ICP/MS) - Total Recoverable**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	0.0350		0.00400	mg/L		07/22/22 11:45	07/23/22 09:12	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 09:12	1

# QC Sample Results

Client: Parkhill Smith & Cooper Inc.  
 Project/Site: Ector County Utility District

Job ID: 880-17149-1

## Method: 200.8 - Metals (ICP/MS)

**Lab Sample ID: MB 860-62070/1-A**  
**Matrix: Drinking Water**  
**Analysis Batch: 62305**

**Client Sample ID: Method Blank**  
**Prep Type: Total Recoverable**  
**Prep Batch: 62070**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Copper	<0.00400	U	0.00400	mg/L		07/22/22 11:45	07/23/22 07:58	1
Lead	<0.00200	U	0.00200	mg/L		07/22/22 11:45	07/23/22 07:58	1

**Lab Sample ID: LCSD 860-62070/3-A**  
**Matrix: Drinking Water**  
**Analysis Batch: 62305**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total Recoverable**  
**Prep Batch: 62070**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Copper	0.100	0.09865		mg/L		99	85 - 115	1	20
Lead	0.100	0.09755		mg/L		98	85 - 115	1	20

**Lab Sample ID: LLCS 860-62070/4-A**  
**Matrix: Drinking Water**  
**Analysis Batch: 62305**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total Recoverable**  
**Prep Batch: 62070**

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Copper	0.00400	0.003622	J	mg/L		91	50 - 150
Lead	0.00200	0.001727	J	mg/L		86	50 - 150

**Lab Sample ID: 880-17149-2 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 62305**

**Client Sample ID: 1735 Jaynita Ave**  
**Prep Type: Total Recoverable**  
**Prep Batch: 62070**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Copper	0.0326		0.100	0.1313		mg/L		99	70 - 130
Lead	<0.00200	U	0.100	0.1028		mg/L		103	70 - 130

# QC Association Summary

Client: Parkhill Smith & Cooper Inc.  
 Project/Site: Ector County Utility District

Job ID: 880-17149-1

## Metals

### Prep Batch: 62070

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
880-17149-1	6033 W. Nixon	Total Recoverable	Drinking Water	200.8	
880-17149-2	1735 Jaynita Ave	Total Recoverable	Drinking Water	200.8	
880-17149-3	W. Swan Rd	Total Recoverable	Drinking Water	200.8	
880-17149-4	5725 W. Turner St.	Total Recoverable	Drinking Water	200.8	
880-17149-5	10464 W. 13th	Total Recoverable	Drinking Water	200.8	
880-17149-6	713 S. Warbonnet Ave.	Total Recoverable	Drinking Water	200.8	
880-17149-7	12371 W. April St.	Total Recoverable	Drinking Water	200.8	
880-17149-8	7172 W. Verita St.	Total Recoverable	Drinking Water	200.8	
880-17149-9	1039 W. Moss Ave	Total Recoverable	Drinking Water	200.8	
MB 860-62070/1-A	Method Blank	Total Recoverable	Drinking Water	200.8	
LCSD 860-62070/3-A	Lab Control Sample Dup	Total Recoverable	Drinking Water	200.8	
LLCS 860-62070/4-A	Lab Control Sample	Total Recoverable	Drinking Water	200.8	
880-17149-2 MS	1735 Jaynita Ave	Total Recoverable	Drinking Water	200.8	

### Analysis Batch: 62305

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
880-17149-1	6033 W. Nixon	Total Recoverable	Drinking Water	200.8	62070
880-17149-2	1735 Jaynita Ave	Total Recoverable	Drinking Water	200.8	62070
880-17149-3	W. Swan Rd	Total Recoverable	Drinking Water	200.8	62070
880-17149-4	5725 W. Turner St.	Total Recoverable	Drinking Water	200.8	62070
880-17149-5	10464 W. 13th	Total Recoverable	Drinking Water	200.8	62070
880-17149-6	713 S. Warbonnet Ave.	Total Recoverable	Drinking Water	200.8	62070
880-17149-7	12371 W. April St.	Total Recoverable	Drinking Water	200.8	62070
880-17149-8	7172 W. Verita St.	Total Recoverable	Drinking Water	200.8	62070
880-17149-9	1039 W. Moss Ave	Total Recoverable	Drinking Water	200.8	62070
MB 860-62070/1-A	Method Blank	Total Recoverable	Drinking Water	200.8	62070
LCSD 860-62070/3-A	Lab Control Sample Dup	Total Recoverable	Drinking Water	200.8	62070
LLCS 860-62070/4-A	Lab Control Sample	Total Recoverable	Drinking Water	200.8	62070
880-17149-2 MS	1735 Jaynita Ave	Total Recoverable	Drinking Water	200.8	62070



# Lab Chronicle

Client: Parkhill Smith & Cooper Inc.  
 Project/Site: Ector County Utility District

Job ID: 880-17149-1

**Client Sample ID: 6033 W. Nixon**

**Lab Sample ID: 880-17149-1**

Date Collected: 07/20/22 07:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 08:44	SHZ	XEN STF

**Client Sample ID: 1735 Jaynita Ave**

**Lab Sample ID: 880-17149-2**

Date Collected: 07/20/22 07:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 08:46	SHZ	XEN STF

**Client Sample ID: W. Swan Rd**

**Lab Sample ID: 880-17149-3**

Date Collected: 07/19/22 16:30

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 08:57	SHZ	XEN STF

**Client Sample ID: 5725 W. Turner St.**

**Lab Sample ID: 880-17149-4**

Date Collected: 07/19/22 11:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 08:59	SHZ	XEN STF

**Client Sample ID: 10464 W. 13th**

**Lab Sample ID: 880-17149-5**

Date Collected: 07/19/22 06:50

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 09:02	SHZ	XEN STF

**Client Sample ID: 713 S. Warbonnet Ave.**

**Lab Sample ID: 880-17149-6**

Date Collected: 07/19/22 12:30

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 09:04	SHZ	XEN STF

# Lab Chronicle

Client: Parkhill Smith & Cooper Inc.  
 Project/Site: Ector County Utility District

Job ID: 880-17149-1

**Client Sample ID: 12371 W. April St.**

**Lab Sample ID: 880-17149-7**

Date Collected: 07/19/22 09:00

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 09:07	SHZ	XEN STF

**Client Sample ID: 7172 W. Verita St.**

**Lab Sample ID: 880-17149-8**

Date Collected: 07/19/22 09:30

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 09:09	SHZ	XEN STF

**Client Sample ID: 1039 W. Moss Ave**

**Lab Sample ID: 880-17149-9**

Date Collected: 07/20/22 11:35

Matrix: Drinking Water

Date Received: 07/20/22 13:20

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total Recoverable	Prep	200.8			50 mL	50 mL	62070	07/22/22 11:45	MD	XEN STF
Total Recoverable	Analysis	200.8		1			62305	07/23/22 09:12	SHZ	XEN STF

**Laboratory References:**

XEN STF = Eurofins Houston, 4147 Greenbriar Dr, Stafford, TX 77477, TEL (281)240-4200

# Accreditation/Certification Summary

Client: Parkhill Smith & Cooper Inc.  
Project/Site: Ector County Utility District

Job ID: 880-17149-1

## Laboratory: Eurofins Houston

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Texas	NELAP	T104704215-22-47	06-30-23

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

# Method Summary

Client: Parkhill Smith & Cooper Inc.  
Project/Site: Ector County Utility District

Job ID: 880-17149-1

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	XEN STF
200.8	Preparation, Total Recoverable Metals	EPA	XEN STF

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

XEN STF = Eurofins Houston, 4147 Greenbriar Dr, Stafford, TX 77477, TEL (281)240-4200



# Sample Summary

Client: Parkhill Smith & Cooper Inc.  
Project/Site: Ector County Utility District

Job ID: 880-17149-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
880-17149-1	6033 W. Nixon	Drinking Water	07/20/22 07:00	07/20/22 13:20
880-17149-2	1735 Jaynita Ave	Drinking Water	07/20/22 07:00	07/20/22 13:20
880-17149-3	W. Swan Rd	Drinking Water	07/19/22 16:30	07/20/22 13:20
880-17149-4	5725 W. Turner St.	Drinking Water	07/19/22 11:00	07/20/22 13:20
880-17149-5	10464 W. 13th	Drinking Water	07/19/22 06:50	07/20/22 13:20
880-17149-6	713 S. Warbonnet Ave.	Drinking Water	07/19/22 12:30	07/20/22 13:20
880-17149-7	12371 W. April St.	Drinking Water	07/19/22 09:00	07/20/22 13:20
880-17149-8	7172 W. Verita St.	Drinking Water	07/19/22 09:30	07/20/22 13:20
880-17149-9	1039 W. Moss Ave	Drinking Water	07/20/22 11:35	07/20/22 13:20

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13



PWS Name: **FEEDS COUNTY UTILITY DISTRICT**  
 PWS ID #: **TX 0680235**  
 PWS Address: **1039 N. MOSS AVE.**  
 PWS Contact Name: **DAVE CLAUSEN**  
 PWS Contact Phone: **432-556-2056**  
 Laboratory Name: **EUROFINS XEROX LLC - HOUSTON**  
 Laboratory Address: **4147 GREENBRIAR DR STAFFORD TX 77477**  
 TCEQ Laboratory ID #: **TD4704215**  
 Laboratory Contact Name: **JULIAN MARTINEZ**  
 Laboratory Contact Phone #: **281-240-4200**

PWS Tap Sample Checklist (V)  
 Samples filled to 1 Liter line  
 Samples taken from an inside sink  
 Samples collected from cold water tap(s)  
 Sinks were unused for 6 hours prior to collector  
 Sample containers w/ labels and homeowner instructions  
 Samples delivered to lab within 14 days of collection

Sample Type (V)  Compliance  Non-Compliance  
 I acknowledge that the information on this form is true and correct and sites selected for sampling follow the approved TCEQ Form 2067 and the PWS Monitoring Plan. Falsification of this form or tampering with water samples is a crime punishable under state and/or federal law. (Texas Penal Code, Title 8, Chapter 37.10)  
 Name of Authorized PWS Representative (Print): **Jose A. Hernandez**  
 Signature: **Jose A. Hernandez**  
 Date: **7-20-22**

Facility ID (e.g. DSO1, PCUC001)	Sample Point ID (e.g. E1C3)	Sample Location (Address of sample point - 133 Main St, Kitchen sink)	Sample Faucet Last Used		Sample Collected		Replacement (S)	Original Sample ID #	Original Collection Date (MMDDYY)	Lab Sample ID #	Lab Preservation		Sample Rejection Code (if applicable)
			Date (MMDDYY)	Time - 24 HR (HHMM)	Date (MMDDYY)	Time - 24 HR (HHMM)					Date (MMDDYY)	Time - 24 HR (HHMM)	
DS01	LCR001	6033 W. N. YON	7-19-22	6:30 P.M.	7-20-22	07:00				880-17149-1	072022	1430	
DS01	LCR002	1735 Jovita, Ave.	7-19-22	9:30 P.M.	7-20-22	07:00				880-17149-2	072022	1430	
DS01	LCR003	W. Sudan Rd	7-19-22	9:30 P.M.	7-19-22	14:30				880-17149-3	072022	1430	
DS01	LCR004	5725 W. Turner St.	7-18-22	10:00 P.M.	7-19-22	11:00				880-17149-4	072022	1430	
DS01	LCR005	10464 W. 13th.	7-18-22	9:30 P.M.	7-19-22	6:50				880-17149-5	072022	1430	
DS01	LCR006	713 S. Wackermet Ave.	7-18-22	10:30 P.M.	7-19-22	12:30				880-17149-6	072022	1430	
DS01	LCR007	12371 W. April St.	7-18-22	8:30 P.M.	7-19-22	19:30				880-17149-7	072022	1430	
DS01	LCR008	7122 W. Venita St.	7-18-22	10:30 P.M.	7-19-22	9:30				880-17149-8	072022	1430	
DS01	LCR009	1039 W. Moss Ave.	7-19-22	8:00 A.M.	7-20-22	11:35				880-17149-9	072022	1430	

Chain of Custody  
 Requisitioned By (Signature): **Jose A. Hernandez**  
 Date/Time: **7-20-22/1:30**  
 Received by Counter (Signature): **Julian Martinez**  
 Date/Time: **7/20/22**  
 Date/Time: **7/20/22**  
 Date/Time: **7/20/22**

## Login Sample Receipt Checklist

Client: Parkhill Smith & Cooper Inc.

Job Number: 880-17149-1

**Login Number: 17149**

**List Number: 1**

**Creator: Rodriguez, Leticia**

**List Source: Eurofins Midland**

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	



## Login Sample Receipt Checklist

Client: Parkhill Smith & Cooper Inc.

Job Number: 880-17149-1

**Login Number: 17149**

**List Number: 2**

**Creator: Milone, Jeancarlo**

**List Source: Eurofins Houston**

**List Creation: 07/21/22 01:47 PM**

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	

